

CERTIFICATE OF ANALYSIS

Product Name H-Tyr-Cys-Phe-Ala-Trp-Lys-Thr-Phe-Cys-OH peptide, Cat.#: 314828 (TFA removed)

MTFA30551-1007 Lot No.

YC*FAWKTFC* (Acetate Salt) Sequence

100%H2O **Dissolution condition**

9AA Length

Modification acetate salt: disulfide bond

Molecular Weight (MW) 1166.39

-20°C Storage

Test Items	Specifications	Results
MW by MS	1167.10	Conforms
Purity by HPLC	>95%	97.398%
Peptide Content	N/A	N/A
Moisture content	N/A	N/A
Acetic acid content	N/A	N/A
Appearance	White to off-white lyophilized powder	Conforms
Quantity	10mg	5.0mg*2

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:

Quality Assurance Department

Nov/03/2023



Sample Information

Order ID :MTFA30551

Name :H-Tyr-Cys-Phe-Ala-Trp-Lys-Thr-Phe-Cys-OH peptide

Sequence :YC*FAWKTFC* (Acetate Salt)

Lot. No :MTFA30551-1007

Pump A :0.1%Trifluoroacetic in 100% water Pump B :0.1%Trifluoroacetic in 100% acetonitrile

Total Flow :1ml/min Wavelength :214nm

Analytial column type :Phenomenex Luna® 5 μm C18(2) (4.6*250mm*5 μm)


Dissolution method :100%H2O

Inj. Volume : 15ul Time Module Value Action 0.01 **Pumps** B.Conc 20 30.00 Pumps **B.Conc** 50 33.00 Pumps B.Conc 100 38.00 100 Pumps B.Conc 20 40.00 Pumps B.Conc

50.00 Controller Stop

Chromatogram

mV




Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	15.178	4459	393	0.048
2	15.696	8900	1218	0.096
3	16.642	169201	3810	1.824
4	17.475	9035826	1159100	97.398
5	17.779	58869	1348	0.635
Total		9277255	1165868	100.000



MS Spectrum



Sample Information

Dissolution method	:15% ACN+85% H2O	Interface	:ESI	Prerod Bias	:+4.5kv
Date Acquired	:2023/10/16 19:15:59	Nebulizing Gas Flow	:1.50L/min	Detector	:-0.2kv
Injection Volume	:0.2ul	CDL Temp	:250°C	T.Flow	:0.2ml/min
Block Temp	:200	CDL Volt	:0v	B.conc	:50% H2O/50% MEOH

Order ID

:MTFA30551 :H-Tyr-Cys-Phe-Ala-Trp-Lys-Thr-Phe-Cys-OH peptide :YC*FAWKTFC* (Acetate Salt) :MTFA30551-1007 Name

Sequence Lot.No

Theoretical :1166.39 Observed :1167.10