
CERTIFICATE OF ANALYSIS

Product Name	H-Ser-Pro-OH peptide, Cat.#: 304454
Lot No.	MTFB20434-260104
Sequence	SP
Dissolution condition	100%H2O
Modification	N/A
Molecular Weight (MW)	202.20
Storage	-20°C

Test Items	Specifications	Results
MW by MS	202.10	Conforms
Purity by HPLC	>95%	95.605%
Endotoxin Content	N/A	N/A
TFA solubility	N/A	N/A
Moisture content	N/A	N/A
Acetic acid content	N/A	N/A
Appearance	White to off-white lyophilized powder	Conforms
Quantity	30mg	5.0mg*6

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:



Quality Assurance Department

Jan/27/2026




Sample Information


Order ID	:MTFB20434		
Name	:N/A		
Sequence	:SP		
Lot. No	:MTFB20434-260104		
Pump A	:0.1%Trifluoroacetic in 100% water		
Pump B	:0.1%Trifluoroacetic in 100% acetonitrile		
Total Flow	:1ml/min		
Wavelength	:214nm		
Analytical column type	:Kromasil 100-5-C18 (4.6*250mm*5 μm)		
Dissolution method	:100%H2O		
Inj. Volume	: 50ul		
Time	Module	Action	Value
0.01	Pumps	B.Conc	0.1
30.00	Pumps	B.Conc	30
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	0.1
50.00	Controller	Stop	

Chromatogram

mV


Peak Table
Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	7.374	4736455	239305	95.605
2	7.779	43147	8239	0.871
3	9.762	174594	9678	3.524
Total		4954196	257223	100.000



Sample Information

Dissolution method	:15%ACN+85%H2O	Interface	:ESI	Prerod Bias	:+4.5kv
Date Acquired	:2026/01/23 13:11:54	Nebulizing Gas Flow	:1.50L/min	Detector	:-0.2kv
Injection Volume	:0.2ul	CDL Temp	:250°C	T.Flow	:0.2ml/min
Block Temp	:200	CDL Volt	:0v	B.conc	:50%H2O/50%MEOH

Order ID	:MTFB20434
Name	:N/A
Sequence	:SP
Lot.No	:MTFB20434-260104
Theoretical	:202.20
Observed	:202.10